AVANTIK

Certificate of Analysis

7/2/2025

19 Chapin Road - Building C, Pine Brook, NJ 07058

Product No.: GR2030-1000

Description: Avantik prefilled 10% Neutral Buffered Formalin, 1000 ml

Lot No.: 43738

Date of Manufacturing: 6/25/2025

Expiration Date: 6/1/2027

10% Formalin Lot #: 5F012

Specification	Requirements	Results
Appearance	Clear, colorless, no observable particulate	Conforms
Formaldehyde Concentration	9.50 -10.50% (v/v)	9.85% (v/v)
pH @ 25°C	6.90 - 7.10	7.04

10% Formalin Lot #: 5F015

Specification	Requirements	Results
Appearance	Clear, colorless, no observable particulate	Conforms
Formaldehyde Concentration	9.50 -10.50% (v/v)	10.00% (v/v)
pH @ 25°C	6.90 - 7.10	7.01

Country of Origin: United States of America

Latex Free: This product contains no latex

This is a representation of an electronic record that was signed electronically and this page is the manifestation of the electronic signatures.

Signatory Table

Action Name	User Name	Title	Signature Date
Send for Approval	Arvind Bandi	PRODUCT MARKETING INTERN	17-Jul-2025 16:01
Approve	Tim Lau	DIGITAL MARKETING INTERN	17-Jul-2025 16:18

* Dates are displayed according to the system time zone: (GMT-04:00) Eastern Daylight Time (America/New_York)