AVANTIK

Certificate of Analysis

19 Chapin Road - Building C, Pine Brook, NJ 07058

11/5/2025

Product No.: GR2030-1000

Description: Avantik prefilled 10% Neutral Buffered Formalin, 1000 ml

Lot No.: 46279

Date of Manufacturing: 11/3/2025

Expiration Date: 10/1/2027

10% Formalin Lot #: 5J015

Specification	Requirements	Results
Appearance	Clear, colorless, no observable particulate	Conforms
Formaldehyde Concentration	9.50 -10.50% (v/v)	9.85% (v/v)
pH @ 25°C	6.90 - 7.10	7.03

Country of Origin: United States of America

Latex Free: This product contains no latex

Electronically Generated Certificate - Valid Without Signature

This is a representation of an electronic record that was signed electronically and this page is the manifestation of the electronic signatures.

Signatory Table

Action Name	User Name	Title	Signature Date
Send for Approval	Arvind Bandi	PRODUCT MARKETING INTERN	09-Dec-2025 09:36
Approve	Jeremy Schriver	DIGITAL COMMUNICATIONS MANAGER	10-Dec-2025 15:58

* Dates are displayed according to the system time zone: (GMT-05:00) Eastern Standard Time (America/New_York)