AVANTIK

Certificate of Analysis

19 Chapin Road - Building C, Pine Brook, NJ 07058

Product No.: GR2030-1000

Description: Avantik Formalin Prefills-1000 mL(12/cs)

Lot No.: 49857

Date of Manufacturing: 6/4/2026

Expiration Date: 6/1/2028

10% Formalin Lot #: 6F001

Specification	Requirements	Results
Appearance	Clear, colorless, no observable particulate	Conforms
Formaldehyde Concentration	9.50 -10.50% (v/v)	10.00% (v/v)
pH @ 25°C	6.90 - 7.10	7.06

10% Formalin Lot #: 6F002

Specification	Requirements	Results
Appearance	Clear, colorless, no observable particulate	Conforms
Formaldehyde Concentration	9.50 -10.50% (v/v)	10.00% (v/v)
pH @ 25°C	6.90 - 7.10	7.01

Country of Origin: United States of America

Latex Free: This product contains no latex

Electronically Generated Certificate - Valid Without Signature

This is a representation of an electronic record that was signed electronically and this page is the manifestation of the electronic signatures.

Signatory Table

Action Name	User Name	Title	Signature Date
Send for Approval	Arvind Bandi	Associate Portfolio Marketing Manager	18-Jun-2026 13:19
Approve	Jeremy Schriver	Senior Manager, Digital Communications	22-Jun-2026 15:59

* Dates are displayed according to the system time zone: (GMT-04:00) Eastern Daylight Time (America/New_York)