Certificate of Analysis

19 Chapin Road - Building C, Pine Brook, NJ 07058

10/15/2026

Product No.: GR2030-20

Description: Avantik prefilled 10% Neutral Buffered Formalin, 20ml

Lot No.: 51483

Date of Manufacturing: 10/14/2026

Expiration Date: 10/1/2028

10% Formalin Lot #: 6J003

Specification	Requirements	Results
Appearance	Clear, colorless, no observable particulate	Conforms
Formaldehyde Concentration	9.50 -10.50% (v/v)	10.31% (v/v)
pH @ 25°C	6.90 - 7.10	7.00

10% Formalin Lot #: 6J004

Specification	Requirements	Results
Appearance	Clear, colorless, no observable particulate	Conforms
Formaldehyde Concentration	9.50 -10.50% (v/v)	10.31% (v/v)
pH @ 25°C	6.90 - 7.10	6.97

Country of Origin: United States of America

Electronically Generated Certificate - Valid Without Signature

Latex Free: This product contains no latex

This is a representation of an electronic record that was signed electronically and this page is the manifestation of the electronic signatures.

Signatory Table

Action Name	User Name	Title	Signature Date
Send for Approval	Jacob Klaff	Associate Product Portfolio Manager	22-Oct-2026 09:58
Approve	Samantha Pignatello	Digital Marketing Associate/Marketing Coordinator	27-Oct-2026 11:38

* Dates are displayed according to the system time zone: (GMT-04:00) Eastern Daylight Time (America/New_York)