AVANTIK
19 Chapin Road - Building C, Pine Brook, NJ 07058
800.783.9424 | avantik-us.com

CERTIFICATE OF ANALYSIS

GR2056-BL - Tissue Marking Dye, Avantik, Blue, 2oz, 12/cs (MTO)

Lot Number: 197366

Expiration Date: 2026-May-31

COMPONENTS

Test	Specifications	Results
Color	Blue	Blue
Viscosity @ 20 - 30 C (cps)	>= 380 and <= 580	577
pH	>= 7.0 and <= 10.0	7.1

For In Vitro Use. May be suitable for other laboratory applications. Consult standard histology and pathology texts and operating procedures for details.

Quality Controll - This report is electronically produced and is valid without signature.

Source: SOL-9007A/197562
Form: WM2103_V5

This is a representation of an electronic record that was signed electronically and this page is the manifestation of the electronic signatures.

Signatory Table

Action Name	User Name	Title	Signature Date
Send for Approval	James Andrew Dunn	SENIOR PRODUCT MARKETING MANAGER	28-May-2024 17:45
Approve	Jeremy Schriver	DIGITAL COMMUNICATIONS MANAGER	29-May-2024 10:15

* Dates are displayed according to the system time zone: (GMT-04:00) Eastern Daylight Time (America/New_York)